

## EC DECLARATION OF CONFORMITY

We

**J. Morita Manufacturing Corporation**  
680 Higashi-hama Minami-cho, Fushimi-ku, Kyoto, 612-8533 Japan

declare under own responsibility, that the product:

Kind of product :	Dental root canal measuring and Treatment units
Model :	DP-XX
Type:	RCM-EX, TR-EX
Medical product Class :	IIa Rule 9

is in compliance with the European Directive :

**93/42/EEC**

**" Council Directive of 14 June 1993 on the approximation of the laws of  
the Member States concerning medical devices."**

The compliance with requirements of EC Directive 93/42/EEC Annexes II has been approved by the following notified body.


Notified Body: TÜV Rheinland LGA Products GmbH  
Tillystrasse 2, 90431 Nuremberg, Germany  
Identification No.: 0197

EU authorized representative:	Medical Technology Fremd Consulting Altenhofstrasse 80, 66386 St. Ingbert, Germany
-------------------------------	--

This declaration of conformity is released to each product release document.

Kyoto, Japan  
2014-07-16

(Place, date)

  
Ogino, Sekiya  
Executive Director