

# J.MORITA MFG.CORP.

680 Higashihama Minami-cho, Fushimi-ku, Kyoto 612-8533, Japan

## EC DECLARATION OF CONFORMITY

We

**J. Morita Manufacturing Corporation**  
680 Higashihama Minami-cho, Fushimi-ku, Kyoto, 612-8533 Japan

declare under own responsibility, that the product:

Kind of product :	Endodontic Treatment Motorized Handpiece
Model :	TR-CM
Medical product Class :	IIa Rule 9

is in compliance with the European Directive :

**93/42/EEC**

“ Council Directive of 14 June 1993 on the approximation of the laws of  
the Member States concerning medical devices.”

The compliance with requirements of EC Directive 93/42/EEC Annex II has been approved by the following notified body.


Notified Body: TÜV Rheinland LGA Products GmbH  
Tillystrasse 2, 90431 Nuremberg, Germany  
Identification No.: 0197

EU authorized representative:	Medical Technology Promedt Consulting Altenhofstrasse 80, 66386 St. Ingbert, Germany
-------------------------------	--

This declaration of conformity is related to each product release document.

Kyoto, Japan  
2010-12-06

(Place, date)



Ogino, Sekiya  
Executive Director

TEL +81-75-611-2141 / FAX +81-75-605-2353