



J. MORITA MFG. CORP.

680 HIGASHIHAMA MINAMI-CHO, FUSHIMI-KU, KYOTO 612-8533, JAPAN

EC DECLARATION OF CONFORMITY

We

J. Morita Manufacturing Corporation
680 Higashihama Minami-cho, Fushimi-ku, Kyoto, 612-8533 Japan
declare under own responsibility, that the product:

Kind of product : Dental Handpiece Maintenance Unit

Model : HIM-1

Type : CE

is in compliance with the European Directive :

73/23/EEC

“ Council Directive of 19 February 1973 on the harmonization of the laws of the Member States relating to electrical designed for use within certain voltage limits.”

The full compliance with the standards listed below proves the conformity with the essential requirements of the above-mentioned EC Directive as separately recorded in our technical documents..

Standard number


EN 60601-1:1988 + A1:1991+ A2:1995
EN 60601-1-2:2001

The accredited test laboratory for the standards is:
VDE Testing and Certification Institute Section FG12
Merianst. 28, 63069 Offenbach, GERMANY

EU authorized representative : Medical Technology Promedt Consulting
Altenhofstrasse 80,
66386 St. Ingbert, Germany

Kyoto, Japan, 2007-1-15

(Place, date)



Sekiya Ogino
Managing Director